PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel

and hereby declares that the Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Chemical, Dimensional, Electrical, Fluid Quantities, Mass, Force, and Weighing Devices, Mechanical, Optical, Thermodynamic, and Time and Frequency Calibration

(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:	*Issue Date:*	*Expiration Date:*
Decembeer 21, 2011	Auust 06, 2024	November 30, 2026

Revision Date:	*Extenstion Date*	*Accreditation No.:*	*Certificate No.:*
January 20, 2026	03/30/2027	72305	L24-601-R2

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	pH Meters	4 pH	0.02 pH	Buffer Standards	ASTM D1293 W-003	F1, F2, F3	F, O
Chemical	pH Meters	7.0 pH	0.02 pH	Buffer Standards	ASTM D1293 W-003	F1, F2, F3	F, O
Chemical	pH Meters	10.0 pH	0.03 pH	Buffer Standards	ASTM D1293 W-003	F1, F2, F3	F, O
Chemical	Conductivity Meters	10 µS to 100 mS	0.1 % of reading	Conductivity standards	W-006	F1, F3	F, O
Chemical	Conductivity Meters	0.054 µS to 100 mS	0.1 % of reading	Conductivity standards	W-006	F1, F3	F, O
Chemical	CO2 Analyzers	Up to 2.5%	0.16 % of reading	Gas Standards G-100	W-018	F1, F3	F, O
Chemical	CO2 Analyzers	2.6 % to 5.9%	0.17 % of reading	Gas Standards G-100	W-018	F1, F3	F, O
Chemical	CO2 Analyzers	6 % to 10 %	0.2 % of reading	Gas Standards G-100	W-018	F1, F3	F, O
Electrical	Electrical Simulation of pH Meters, Ion Meters & Titrators (Voltage)	-3 000 mV to 3 000 mV	(0.000 58 + 4 x 10^{-5}V) mV	Glass electrode	W-005 W-003	F1, F3	F, O
Electrical	Conductivity Meters	0.1 µS to 100 mS	0.1 % of reading	Resistance Decade	W-006	F1, F3	F, O
Electrical	Electrical Simulation of Temperature Measuring Indicators (Resistance & Voltage)	-200 °C to 200 °C	0.03 °C	IET RS201W Precision Resistance Substitute	ASTM E2593 W-003	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	1 mg to 500 mg	(2.6 x 10^{-3} + 5.36 x 10-^{6}Wt) mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	500 mg to 1 g	(4.5 x 10^{-3} + 1.44 x 10^{-6}Wt) mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Mass, Force, and Weighing Devices	Analytical Balances	1 g to 60 g	$(1.29 \times 10^{-2} + 3.94 \times 10^{-7}Wt)$ mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	60 g to 300 g	$(1.14 \times 10^{-1} + 1.05 \times 10^{-7}Wt)$ mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	300 g to 500 g	$(1.23 \times 10^{-1} + 7.57 \times 10^{-8}Wt)$ mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	500 g to 1 000 g	$(1.5 \times 10^{-3} + 1.45 \times 10^{-8}Wt)$ mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	1 000 g to 8 000 g	$(1.16 \times 10^{-2} + 5.96 \times 10\text{-}^{8}Wt)$ mg	OIML Class E2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Mass, Force, and Weighing Devices	Analytical Balances	8 001 g to 15 000 g	$(1.18 \times 10^{-2} + 1.12 \times 10\text{-}6Wt)$ g	OIML Class F2 Weights	OIML R 76-1 USP41 EURAMET CG-18 W-001	F1, F2, F3	F, O
Dimensional	Titrator (Spindle Travel)	0.015 mm to 100 mm	0.004 6 mm	Digital Indicator	W-005	F1, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	≤ 1 µL to <10 µL	0.027 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	10 µL to < 20 µL	0.044 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	20 µL to < 50 µL	0.045 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	50 µL to < 100 µL	0.046 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	100 µL to < 200 µL	0.05 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	200 µL to < 500 µL	0.18 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	500 µL to < 1ml	0.34 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	1 mL to < 2ml	0.32 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	2 mL to < 5ml	0.47 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	5 mL to < 10ml	0.49 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	10 mL to < 20ml	0.49 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	20 mL to <50 ml	0.63 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	50 mL to < 100ml	2.5 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Fluid Quantities	Pipettes, Burettes, Dispensers, Dilutors Syringes	100 mL	15 µL	Analytical Balances	ISO 8655 W-002-(Burette) W-004-(Pipettes) W-008-(Dispenser) W-009 (Dilutor) W-019 (Syringes) Gravimetric Method	F1, F2, F3	F, O
Mechanical	Pressure chambers (Autoclaves and Vacuum Ovens)	-12 psi to 100 psi	0.27 psi	Fluke Pressure Gauge, Fluke 718 100G Pressure Calibration 100 PSIG	EURAMET CG-17 EURAMET CG-11 W-016	F1, F2, F3	F, O
Mechanical	Pressure Gauges	-12 psi to 100 psi	0.27 psi	Fluke Pressure Gauge	EURAMET CG-17	F1, F2	F, O
Mechanical	Atmospheric Pressure Gauges	600 mBar to 1 030 mBar	0.8 mBar	Lutron MHB-382	W-016	F1, F3	F, O
Mechanical	Differential pressure Gauges	0 Pa to 200 Pa	0.75 Pa	Digital monometer	EURAMET CG-17 W-016	F1, F2, F3	F, O
Mechanical	Differential Pressure Gauges	201 Pa to 1 000 Pa	1 Pa	Digital monometer	EURAMET CG-17 W-016	F1, F2, F3	F, O
Mechanical	Differential Pressure Gauges	1 001 Pa to 5 000 Pa	3 Pa	Digital monometer	EURAMET CG-17 W-016	F1, F2, F3	F, O
Mechanical	Differential Pressure Gauges	5 001 Pa to 9 000 Pa	5 Pa	Digital monometer	EURAMET CG-17 W-016	F1, F2, F3	F, O
Optical	Spectrophotometer, Photometer (Photometer)	Up to 8 ABS	0.004 6 ABS	Starna reference filters	W-014	F1, F3	F, O
Optical	Spectrophotometer, Photometer (Wavelength Bandwidth)	190 nm to 800 nm	0.11 nm	Starna reference filters	W-014	F1, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Thermodynamic	Climatic Chamber – calibration and uniformity (Oven Incubator, Cooler, Freezer)	-80°C to 300°C	0.07°C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Climatic Chamber – calibration and uniformity (Oven Incubator, Cooler, Freezer)	300°C to 400°C	0.08 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Climatic Chamber – calibration and uniformity (Oven Incubator, Cooler, Freezer)	400 °C to 500 °C	0.07 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Climatic Chamber – calibration and uniformity (Oven Incubator, Cooler, Freezer)	500 °C to 1 000 °C	0.75 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Temperature Probes (RTD, Thermocouple, Liquid Thermometer)	-197 °C to 200 °C (Liquid Nitrogen)	0.03 °C	Bath/Dry Bath: Fluke 7103, Reference calibration probe Fluke 1523, Grant Logger	ASTM E2593 ASTM E77 EURAMET CG-8 W-012	F1, F2, F3	F, O
Thermodynamic	Temperature Probes (RTD, Thermocouple, Liquid Thermometer)	201 °C to 400 °C	0.05 °C	Bath/Dry Bath: Fluke 7103, Reference calibration probe Fluke 1523, Grant Logger	ASTM E2593 ASTM E77 EURAMET CG-8 W-012	F1, F2, F3	F, O
Thermodynamic	Temperature Probes (RTD, Thermocouple, Liquid Thermometer)	401 ° to 550 °C	1 °C	Bath/Dry Bath: Fluke 7103, Reference calibration probe Fluke 1523, Grant Logger	ASTM E2593 ASTM E77 EURAMET CG-8 W-012	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Thermodynamic	Temperature Probes (RTD, Thermocouple, Liquid Thermometer)	551 ° to 800 °C	1.4 °C	Bath/Dry Bath: Fluke 7103, Reference calibration probe Fluke 1523, Grant Logger	ASTM E2593 ASTM E77 EURAMET CG-8 W-012	F1, F2, F3	F, O
Thermodynamic	Temperature Probes (RTD, Thermocouple, Liquid Thermometer)	801 °C to 1 200 °C	2.3 °C	Bath/Dry Bath: Fluke 7103, Reference calibration probe Fluke 1523, Grant Logger	ASTM E2593 ASTM E77 EURAMET CG-8 W-012	F1, F2, F3	F, O
Thermodynamic	Liquid Baths	-20 C to 150 °C	0.03 °C	Fluke 1523, Grant Logger Reference	ASTM E715 EURAMET CG-13 W-011	F1, F2, F3	F, O
Thermodynamic	Precision Temperature Blocks	-80 °C to 200 °C	0.03 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Precision Temperature Blocks	200°C to 300°C	0.04 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Precision Temperature Blocks	300°C to 400°C	0.05 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O
Thermodynamic	Precision Climatic Chamber	0 °C to 50 °C	0.03 °C	ISOTECH / Fluke reference Calibration probe	W-010 Euramet CG-20 DKD-R 5-7 AMS2750	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD

29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY (±) [1]	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Thermodynamic	Oven/Incubator/ Cooler/Freezer and Uniformity Climate Chambers (Humidity)	10 % RH to 95 % RH	1.5 % RH	Grant Humidity Logger	AMS2750 EURAMET CG-20 DKD-R 5-7 W-010	F1, F2, F3	F, O
Thermodynamic	Humidity Sensor with Indicator	10 % RH to 95 % RH	0.7 % RH	Humidity Probes	W-012	F1, F3	F, O
Time and Frequency	Centrifuges, Shakers (Rotational Speed)	10 RPM to 60 000 RPM	1.1 RPM	Optical- Tachometer	W-015	F1, F3	F, O
Time and Frequency	Stopwatch/Timers	Up to 24 hrs	0.6 sec/day	Stopwatch	W-017 NIST 960-12	F1, F2, F3	F, O

Certificate of Accreditation: Supplement

DS Bioanalytics, LTD
29th Omarim Street, Omer, 84965 Israel
Contact Name: Dani Shahar Phone: +97-272-242-9555

Accreditation is granted to the facility to perform the following conformity assessment activities:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.

2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.

3. Location of activity:

Location Code	Location
F	Conformity assessment activity is performed at the CABs fixed facility
O	Conformity assessment activity is performed onsite at the CABs customer location

4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.

5. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.

6. Flex Codes

F0: When no flexibility is identified. There are no changes to items calibrated, characteristics identified or versions of methods except for updating to the most recent version of a standard method after verification.
F1: The laboratory has the capability to introduce a new instrument, quantity, or gauge for an accredited calibration method.
F2: The laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope
F3: The laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope
F4: The laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using the same Calibration Equipment or Reference Standards identified on the scope for the same parameter, component, or analyte identified on the line item of the scope.